

Stavros Kromidas

 WILEY-VCH

More Practical Problem Solving in HPLC



Stavros Kromidas

More Practical Problem Solving in HPLC

with Contributions by
Friedrich Mandel, Jürgen Maier-Rosenkranz and Hans-Joachim Kuss

Translated by
Renate FitzRoy



WILEY-
VCH

WILEY-VCH Verlag GmbH & Co. KGaA

Dr. Stavros Kromidas

Rosenstraße 16
66125 Saarbrücken
Germany
info@kromidas.de

Dr. Friedrich Mandel

Senior Applications Chemist
Mass Spectrometry
Agilent Technologies
Sales & Services GmbH & Co. KG
Hewlett-Packard-Str. 8
76337 Waldbronn
Germany
friedrich_mandel@agilent.com

Dr. Juergen Maier-Rosenkranz

GROM Chromatography GmbH
a GRACE Vydac Division
Etzwiesenstraße 37
72108 Rottenburg-Hailfingen
Germany
Juergen.Maier-Rosenkranz@grace.com

Dr. Hans-Joachim Kuss

Innenstadtklinikum der LMU
Nussbaumstr. 7
80338 München
Germany
Hans-Joachim.Kuss@med.uni-muenchen.de

This book was carefully produced. Nevertheless, authors and publisher do not warrant the information contained therein to be free of errors. Readers are advised to keep in mind that statements, data, illustrations, procedural details or other items may inadvertently be inaccurate.

Library of Congress Card No.: applied for

A catalogue record for this book is available from the British Library.

Bibliographic information published by Die Deutsche Bibliothek

Die Deutsche Bibliothek lists this publication in the Deutsche Nationalbibliografie; detailed bibliographic data is available in the Internet at <http://dnb.ddb.de>

© 2005 WILEY-VCH Verlag GmbH & Co. KGaA, Weinheim

All rights reserved (including those of translation in other languages). No part of this book may be reproduced in any form – by photoprinting, microfilm, or any other means – nor transmitted or translated into machine language without written permission from the publishers. Registered names, trademarks, etc. used in this book, even when not specifically marked as such, are not to be considered unprotected by law.

Printed in the Federal Republic of Germany.

Printed on acid-free paper.

Typesetting K+V Fotosatz, Beerfelden

Printing betz-druck gmbh, Darmstadt

Bookbinding J. Schäffer GmbH & Co. KG, Grünstadt

ISBN 3-527-31113-0

Contents

Preface XV

The Structure of the Book 1

Part 1 (general section) 1

Part 2 (specific questions) 1

In Lieu of an Introduction 3

Chromatography Crossword 4

Across 4

Down 4

An HPLC-Quiz 6

An HPLC Tale 9

The Tale of Peaky and Chromy 9

1 HPLC Tips 11

1.1 Stationary Phases and Columns 11

Tip No.

01 “It improves with age” is a rule that applies to port and sometimes to red wine, but how about your C₁₈ column? 11

02 Optimization via column parameters – what works best? 14

03 Can selectivity always be put down to chemical interactions with the stationary phase? 17

04 A matter of perspective ... Or: Selectivity and peak symmetry of basic compounds using reversed-phase packing materials 19

05 Separation of isomers 21

06 When should I use a “polar” C₁₈ phase? 23

07 Are polar RP-C₁₈ phases more suitable for the separation of polar analytes than non-polar phases? 24

08 What about non-encapped phases – are they a thing of the past? 25

09 How can I separate acids using RP C₁₈? 27

10 The nitrile phase – some like it polar 29

11 The selectivity of RP columns 31

1.2 Buffers, pH Value 33

12 Does it always have to be potassium phosphate? 33

Tip No.

- 13 UV cut-off of buffer solutions 34
- 14 Sources of errors when using buffers 35
- 15 The drawbacks of using buffers 37
- 16 Why is the pH value so important, and what does it do? 40
- 17 Why does the pH value shift even though I am using the correct buffer and the buffer capacity is sufficient? 42
- 18 Changes to the pH value in the eluent: the extent of the shift and the reasons behind it 43
- 19 An unintentional pH shift and its consequences 46
- 20 RP separations in the alkaline medium 49
- 21 Separation of basic and acidic compounds contained in the same sample 51

1.3 Optimization, Peak Homogeneity 53

- 22 The peaks appear too soon – what can be done? 53
- 23 What can I do if the peaks elute late? 55
- 24 Quick optimization of an existing gradient method 60
- 25 Increasing efficiency – often the fast track to success 63
- 26 Additives to the eluent 66
- 27 Separating the unknown – where shall I begin? 69
- 28 Separation of an unknown sample using a reversed-phase C₁₈ column – how do I go about it? 72
- 29 Developing an RP separation – the two-day-method
Part 1: Choice of column and eluent 74
- 30 Developing an RP separation – the two-day method
Part 2: Fine-tuning of the separation 78
- 31 Quick check on peak homogeneity – Part 1 80
- 32 Quick check on peak homogeneity – Part 2 82
- 33 Tied to a standard operating procedure –
how can a bad separation be improved further? 84
- 34 More elaborate measures to check peak homogeneity 86
- 35 First easily digestible tip 91
- 36 Second easily digestible tip 94
- 37 Third easily digestible tip 96

1.4 Troubleshooting 99

- 38 How to approach problems in a systematic manner 99
- 39 Spikes in the chromatogram 101
- 40 Additional peaks in trace analysis separations 103
- 41 What causes a ghost peak? 105
- 42 Ghost peaks in a blank gradient 107
- 43 Strange behaviour of a peak. What could be the cause? 108
- 44 When could one expect a change in the elution order of the peaks? 110
- 45 Tailing in RP HPLC – Part 1: Fast troubleshooting 114
- 46 Tailing in RP HPLC – Part 2: Further causes and time-served cures 116

Tip No.

- 47 Peak deformation and a shift in retention time due to an unsuitable sample solvent 119
- 48 Is flushing with water or acetonitrile sufficient? 123
- 49 Flushing and washing fluids for HPLC apparatus 125
- 50 When does the peak area change? 127
- 51 Reasons for a change in either peak height or peak area, but not in both 129
- 52 Excesses and their pitfalls 131
- 53 Algae, fungi and bacteria in HPLC 132
- 54 Does 40 °C always mean 40 °C? 134
- 55 The most common reason for a lack of reproducibility is a lack of methodological robustness 135
- Have a break ... 138
- Dear Reader 138
- Complete the sentences 139
- “Matching pairs” 140
- Has Peaky remembered his lessons correctly? 141
- 1.5 General HPLC Tips 142**
- 56 What changes can you expect when switching from one HPLC system to another? 142
- 57 What changes can be expected in a chromatogram if the dead volume is larger in one isocratic system than in another? 144
- 58 Contribution of the individual modules of the system to band broadening 146
- 59 How to keep retention times constant while reducing the diameter of the column 148
- 60 Has 3 µm material been developed sufficiently to be used in routine separations? 150
- 61 Miniaturization may be all well and good – but when does it really work and does it make sense in routine separations? 152
- 62 Why is it that peaks appear later with a new column? 154
- 63 Column length, flow and retention times in gradient separations 155
- 64 Column dimensions and gradient separations 159
- 65 What is the difference between dead time and dead volume on the one hand and selectivity and resolution on the other? 161
- 66 Troublesome small peaks 163
- 67 Lowering the detection limit by optimizing the injection 164
- 68 Setting the parameters of an HPLC instrument 167
- 69 The right wavelength – old hat to some, a revelation to others 171
- 70 Characteristics of refraction, fluorescence and conductivity detectors 175

Tip No.

- 71 Does it always have to be HPLC? 177
72 Methanol versus acetonitrile 180
73 Tale of a foursome pub-crawl – can peaks elute before the front? 183

References to HPLC-Tips 185

2 LC-MS Coupling, Micro- and Nano-LC, Quantification 187

*2.1 by Friedrich Mandel, 2.2 by Jürgen Maier-Rosenkranz,
2.3.2–2.3.4 by Hans-Joachim Kuss*

2.1 LC-MS Coupling 187

LC-MS – The one and only universal tool? 187

Tip No.

- 74 Choosing the right LC-MS interface 189
75 Which mobile phases are compatible with LC-MS? 195
76 Phosphate buffers – the exception 197
77 Paired ions 198
78 Using additives to enhance API-electrospray ionization 200
79 How can I enhance sensitivity of detection? 203
80 No linear response and poor dynamic range? 205
81 How much MS^n do I need? 208

Need more help? 210

References to LC-MS 211

Internet addresses of interest for LC-MS coupling 212

2.2 Micro- and Nano-LC 213

A short introduction 213

Tip No.

- 82 Lower efficiency – plate number too low
Part 1: Effects of dead volume in the connecting parts.
Which column diameter should be used with which capillary
diameter? 214
83 Lower efficiency – plate number too low
Part 2: Effects of injection amount and injection volume 215
84 Lower efficiency – plate number too low
Part 3: Impact of flow cell (UV, fluorescence, radio detection) 216
85 No gain in sensitivity: flow cell – path length – S/N 218
86 Fused silica and PEEK capillary connections 219
87 Fast sample loading due to column switching 220

Tip No.	
88	Injection system: full loop injection, partial loop fill injection, timed programmed injection, direct injection 222
89	Protecting the system: cleaning-up guard column, saturation column 223
90	Retention time shift: gradient delay volume, mixing chamber volume, gradient accuracy 224
91	Transferability – downscaling: correct gradients 226
	References to Micro- and Nano-LC 228

2.3 Quantification 229

2.3.1	Practical aspects of quantification in HPLC 229
2.3.1.1	Peak area or peak height? 229
2.3.1.2	What factors have an impact on the peak area? 230
2.3.1.3	Formulae and short statements or comments with respect to the quantification methods 231
2.3.1.4	Examples with actual figures 233
2.3.2	Quantification in Chromatography 236
2.3.2.1	Optimum separation – correct peak acquisition 236
2.3.2.2	Understanding the “mind” of the integration system 240
2.3.2.3	Setting parameters and their effect on peak area and peak height 241
2.3.2.4	Where can mistakes be made? 243
2.3.3	Methods of Quantification 244
2.3.3.1	What is the 100% method? 244
2.3.3.2	What is the external standard method? 244
2.3.3.3	Why use an internal standard? 245
2.3.3.4	In which cases should the additions method be used? 247
2.3.4	Weighted Regression 249
2.3.4.1	What about the <i>F</i> -test? What are the other possibilities? 249
2.3.4.2	How do we weight the individual values? 250
2.3.4.3	How to use Excel for weighted regression 250
2.3.5	Solutions to quantification problems 253

References 253

3 Appendix 261

3.1 Solutions to the Problems 261

3.1.1	Crossword – the solution 261
3.1.2	An HPLC quiz – the solution 261
3.1.3	An HPLC tale with Peaky and Chromy – the solution 264
3.1.4	Complete the sentences 265
3.1.5	“Matching pairs” 266
3.1.6	Did Peaky remember his lessons correctly? 267

3.2	From Theory to Practice – Empirical Formulae, Rules of Thumb and Simple Correlations in Everyday HPLC	269
3.3	Information Resources for Analysis/HPLC	277
3.4	Analytical Chemistry Today	281
3.5	Trends in HPLC	285
3.6	Thoughts About a Dead Horse	290
Index		291